

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 30 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130177</u>	
Name <u>STEVEN C. KEVAN</u>		ID# <u>6585</u>	Date <u>6/25/09</u>
A	Agency <u>ALASKA STATE TROOPERS</u>		Phone # <u>800-240-2019</u>
Instrument Location <u>ST. MARY'S</u>			
Alco S/N <u>X 301550</u>		Target Value <u>.079</u>	High Pressure <u>950</u>
B	Alco Test Value Average <u>.086</u> <u>.081</u>		
	1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco	
Signature <u>[Signature]</u>		<u>7/15/09</u>	
(OVER)			

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 1st day of 20th, 2009.

[Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm breath Test Instrument*

*State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N 130177 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130177

JUNE 25, 2009

OPERATOR'S NAME:  
KEUHAN, STEVE  
OPERATOR'S NUMBER: 6585  
SUBJECT'S LAST NAME:  
VOC TEST  
SUBJECT'S FIRST NAME/MI :  
VOC  
O.L. #: 0  
DEPT/AGENCY: SHV0  
CASE/REPORT: 0-0  
TEST TYPE: T  
ALCO TARGET VALUE: .079 ✓  
ALCO S/N: X301550 ✓

### --- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.77 in  
ALCO TARGET .078 12:29  
BLANK TEST .000 12:30  
INTERNAL STANDARD VERIFIED 12:30  
ALCO TO 29.77 in .000 ✓ 12:30  
BLANK TEST .000 ✓ 12:31  
SUBJECT SAMPLE .000 ✓ 12:32  
BLANK TEST .000 ✓ 12:32  
ALCO TO 29.78 in .061 12:33  
BLANK TEST .000 12:34

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130177

JUNE 25, 2009 ✓  
TIME 12:25

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 29.77 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!""#%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~